

Product Information

REExtract-N-Amp™ Blood PCR Kits

Product Codes **XNABS, XNAB, XNABE, XNABR, XNABRE, and P 8240**

TECHNICAL BULLETIN

Product Description

The REExtract-N-Amp™ Blood PCR Kits contain all of the reagents required to rapidly extract and amplify human genomic DNA from whole blood, whole blood dried on a blood card, and cultured mammalian cells. Briefly, DNA is released by incubating the sample with the Lysis Solution at room temperature for 5 minutes for whole blood, at 55 °C for 15 minutes for blood cards, or at 75 °C for 5 to 10 minutes for cell monolayers. After adding the Neutralization Solution, the extract is ready for PCR.

An aliquot of the neutralized extract is then combined with the REExtract-N-Amp Blood PCR ReadyMix™ and user-provided PCR primers to amplify the target DNA. The REExtract-N-Amp Blood PCR ReadyMix is a 2x reaction mixture containing buffer, salts, dNTPs, and *Taq* polymerase. It also contains the JumpStart™ *Taq* antibody for hot start PCR to enhance specificity and the REDTaq™ dye to allow direct loading of the PCR product onto an agarose gel.

Reagents Provided	Product Code	XNABS 10 preps, 10 PCRs	XNAB 100 preps, 100 PCRs	XNABE 100 preps, 500 PCRs	XNABR 1,000 preps, 1,000 PCRs	XNABRE 1,000 preps, 5,000 PCRs
Lysis Solution for Blood	L 3289	0.3 ml	2.5 ml	2.5 ml	25 ml	25 ml
Neutralization Solution for Blood	N 9784	2 x 1.5 ml	25 ml	25 ml	250 ml	250 ml
REExtract-N-Amp Blood PCR ReadyMix. This is a 2x PCR reaction mix containing buffer, salts, dNTPs, <i>Taq</i> polymerase, REDTaq dye, and JumpStart <i>Taq</i> antibody.	P 8240	0.15 ml	1.2 ml	5 x 1.2 ml	12 ml	5 x 12 ml

Storage

The REExtract-N-Amp Blood PCR Kits can be stored at 2-8 °C on a short-term basis up to 3 weeks. For long-term storage greater than 3 weeks, -20 °C is recommended. Do not store in a "frost-free" freezer.

Reagents and Equipment Required But Not Provided

- Microcentrifuge tubes or multiwell plate for extractions (200 µl minimal volume)
- Punch and cards for dried blood
- Incubator or oven for blood cards (55 °C) or monolayer cells (75 °C)
- Tubes or plate for PCR
- Thermal cycler
- PCR primers
- Water, PCR reagent, Product Code W 1754

Precautions and Disclaimer

The REExtract-N-Amp Blood PCR Kits are for R&D use only, not for drug, household, or other uses. The Lysis Solution is caustic. Avoid contact with skin. Wear gloves, safety glasses, and suitable protective clothing when handling this or any other reagent provided with the kit. Consult the MSDS for information regarding hazards and safe handling practices.

Procedure

All steps are carried out at room temperature unless otherwise noted.

A. DNA extraction from Whole Blood

- 1a. Collect blood into tubes containing EDTA, sodium citrate, or sodium heparin. The best results may be obtained with EDTA or sodium citrate. Mix thoroughly by inversion or rocking.

Note: For non-human sources, collect blood into tripotassium EDTA (Product Code E 0270) at a final concentration of 5 mM to prevent coagulation.

- 2a. Place 20 μ l of the Lysis Solution for Blood into a microcentrifuge tube or well of a multiwell plate for each extraction.
- 3a. Add 10 μ l of blood. Mix thoroughly by vortexing or pipetting.
- 4a. Incubate at room temperature for 5 minutes.
- 5a. Add 180 μ l of the Neutralization Solution for Blood. Mix thoroughly by vortexing or pipetting.
- 6a. Store the neutralized blood extract at 4 °C or use 2 μ l immediately in PCR. Continue with step 7.
Note: DNA is stable in the extract for at least 6 months at 4 °C.

B. DNA extraction from Blood Cards

- 1b. Collect the blood sample on to a collection card (such as Product Code C 2613). Allow to dry completely.
- 2b. Punch a disk (preferably 1/8 inch or 3 mm) from the blood card and place into a microcentrifuge tube. Make sure that the punch contains as much of the blood-stained area as possible.
- 3b. Pipette 20 μ l of the Lysis Solution for Blood onto the blood card punch. Samples can be spun in a microcentrifuge for a few seconds to force the solution into the punch.
- 4b. Incubate at 55 °C for 15 minutes.
- 5b. Add 180 μ l of the Neutralization Solution for Blood. Mix thoroughly by vortexing or pipetting.
- 6b. Store the neutralized blood extract at 4 °C or use 2 μ l immediately in PCR. Continue with step 7.
Note: DNA is stable in the extract for at least 6 months at 4 °C.

C. DNA Extraction from Cultured Mammalian Cells

- 1c. Grow monolayer cells in a multiwell plate until 90 to 95% confluent.
- 2c. Aspirate the medium from the wells using a pipette tip connected to a vacuum system. The medium must be removed completely.
- 3c. Add 20 μ l of the Lysis Solution for Blood to each of the wells.
Note: It is preferred at this point to seal the plate with AlumaSeal™ II (Product Code A 2350) to prevent loss by evaporation during incubation in step 4c. The Alumaseal can be pierced with a pipette tip to add the Neutralization Solution for Blood in step 5c. A new layer of AlumaSeal can be placed over the original layer to reseal the plate for storage.

- 4c. Incubate the plate at 75 °C for 5 to 10 minutes (for a 24 well plate, 5 minutes is recommended to avoid overdrying the samples).
- 5c. Add 180 μ l of the Neutralization Solution for Blood to each of the wells. Mix the samples by pipetting up and down.
- 6c. Store the neutralized cell extract at 4 °C or use 2 μ l immediately in PCR. Continue with step 7.
Note: DNA is stable in the extract for at least 6 months at 4 °C.

PCR amplification

The REExtract-N-Amp Blood PCR ReadyMix contains the JumpStart *Taq* antibody for specific hot start amplification. Therefore, PCR reactions can be assembled at room temperature without premature *Taq* DNA polymerase activity.

Typical final primer concentrations are approximately 0.4 μ M each. The optimal primer concentration and cycling parameters will depend on the system used.

7. Add the following reagents to a thin-walled PCR microcentrifuge tube or plate:

Reagent	Volume
Water, PCR Reagent	x μ l
REExtract-N-Amp Blood PCR ReadyMix	10 μ l
Forward primer	y μ l
Reverse primer	y μ l
Neutralized blood extract	2 μ l
Total volume	20 μ l

Note: The neutralized blood extract may inhibit PCR amplification of products larger than 2 kb. Neutralization Solution B (Product Code N 3910) can be used to overcome this inhibition and allows successful amplification of longer PCR products. Add 1 μ l of Neutralization Solution B to each reaction. Neutralization Solution B is not part of this kit and must be purchased separately.

8. Mix gently.
9. For thermal cyclers without a heated lid, add 20 μ l of mineral oil on top of the mixture in each tube to prevent evaporation.
10. Perform thermal cycling. The amplification parameters should be optimized for individual primers, template, and thermal cycler (see References for guidance).

Common cycling parameters:

Step	Temp.	Time	Cycles
Initial Denaturation	94-96 °C	3 minutes	1
Denaturation	94-96 °C	0.5-1 minutes	30-40
Annealing	45-68 °C	0.5-1 minutes	
Extension	72 °C	1-2 minutes (~1 kb/min)	
Final Extension	72 °C	10 minutes	1
Hold	4 °C	Indefinitely	

11. The amplified DNA can be loaded directly onto an agarose gel after the PCR is completed. It is not necessary to add a separate loading buffer/tracking dye.

Note: PCR products can be purified, if desired, for applications such as sequencing with the GenElute™ PCR Clean-Up Kit (Product Code NA1020).

Related Products	Product Code
PCR 96 well plates	P 3606; Z37,490-3
PCR 384 well plates	P 4606; Z37,491-1
Sealing mats and tape	P 4356; P 4481; P 5356; Z37,493-8
AlumaSeal II	A 2350
EDTA, tripotassium salt dihydrate	E 0270
1/8 inch Punch	P 9742
Collection Card	C 2613
PCR microtubes	Z37,487-3; Z37,496-2; Z37,488-1
Neutralization Solution B	N 3910
Mineral Oil	M 8662
PCR Marker	P 9577
Precast Agarose Gels	P 5597; P 5847; P 6097
TBE Buffer	T 4415; T 6400; T 9525
GenElute PCR Clean-Up Kit	NA1020

References

- Dieffenbach, C.W., and Dveksler, G.S. (Eds.), PCR Primer: A Laboratory Manual, Cold Spring Harbor Laboratory Press, New York (1995). (Product Code Z36,411-8)
- Don, R.H. et al., 'Touchdown' PCR to circumvent spurious priming during gene amplification. *Nucleic Acids Res.*, **19**, 4008 (1991).

- Erlich, H.A. (Ed.), PCR Technology: Principles and Applications for DNA Amplification, Stockton Press, New York (1989). (Product Code P 3551)
- Griffin, H.G., and Griffin, A.M. (Eds.), PCR Technology: Current Innovations, CRC Press, Boca Raton, FL (1994). (Product Code Z35,749-9)
- Innis, M.A., et al., (Eds.), PCR Strategies, Academic Press, New York (1995). (Product Code Z36,445-2)
- Innis, M., et al., (Eds.), PCR Protocols: A Guide to Methods and Applications, Academic Press, San Diego, California (1990). (Product Code P 8177)
- McPherson, M.J. et al., (Eds.), PCR 2: A Practical Approach, IRL Press, New York (1995). (Product Code Z36,238-7)
- Newton, C.R. (Ed.), PCR: Essential Data, John Wiley & Sons, New York (1995). (Product Code Z36,491-6)
- Roux, K.H. Optimization and troubleshooting in PCR. *PCR Methods Appl.*, **4**, 5185-5194 (1995).
- Saiki, R., PCR Technology: Principles and Applications for DNA Amplification, Stockton, New York (1989).

A license under US Patents 4,683,202, 4,683,195, 4,965,188, and 5,075,216 or their foreign counterparts, owned by Roche Molecular Systems, Inc. and F. Hoffmann-La Roche Ltd ("Roche"), has an up-front fee component and a running-royalty component. The purchase price of this product includes limited, nontransferable rights under the running-royalty component to use only this amount of the product to practice the Polymerase Chain Reaction ("PCR") and related processes described in said patents solely for the research and development activities of the purchaser when this product is used in conjunction with a thermal cycler whose use is covered by the up-front fee component. Rights to the up-front fee component must be obtained by the end user in order to have a complete license to use this product in the PCR process. These rights under the up-front fee component may be purchased from Applied Biosystems or obtained by purchasing an Authorized Thermal cycler. No right to perform or offer commercial services of any kind using PCR, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is hereby granted by implication or estoppel. Further information on purchasing licenses to practice the PCR Process may be obtained by contacting the Director of Licensing at Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404 or the Licensing Department at Roche Molecular Systems, Inc., 1145 Atlantic Avenue, Alameda, California 94501.

Troubleshooting Guide

Problem	Cause	Solution
Little or no PCR product is detected.	PCR reaction is inhibited due to contaminants in the blood extract.	Use less extract or dilute the extract with water and repeat PCR. To test for inhibition, include a DNA control and/or add a known amount of template (100-500 copies) into the PCR mixture along with the blood extract.
	A PCR component is missing or degraded.	Run a positive control to insure components are functioning. A checklist is also recommended when assembling reactions.
	Too few cycles are performed.	Increase the number of cycles (5-10 additional cycles at a time).
	The annealing temperature is too high.	Decrease the annealing temperature in 2-4 °C increments.
	The primers are not designed optimally.	Confirm the accuracy of the sequence information. If the primers are less than 22 nucleotides long, try to lengthen the primer to 25-30 nucleotides. If the primer has a GC content of less than 45%, try to redesign the primer with a GC content of 45-60%.
	The denaturation temperature is too high or too low.	Optimize the denaturation temperature by increasing or decreasing the temperature in 1 °C increments.
	The denaturation time is too long or too short.	Optimize the denaturation time by increasing or decreasing the time in 10 second increments.
	The extension time is too short.	Increase the extension time in 1 minute increments, especially for long templates.
	The target template is complex.	In most cases, inherently complex targets are due to unusually high GC content and/or secondary structure. Betaine has been reported to help amplification of high GC content templates at a concentration of 1.0-1.7 M.
Multiple products are seen.	JumpStart <i>Taq</i> antibody is not working correctly.	Do not use DMSO or formamide with REExtract-N-Amp PCR ReadyMix. It can interfere with the enzyme-antibody complex. Other solvents, salts, and extremes in pH or other reaction conditions may reduce the affinity of the JumpStart <i>Taq</i> antibody for the <i>Taq</i> polymerase and thereby compromise its effectiveness.
	Touchdown PCR may be needed.	“Touchdown” PCR significantly improves the specificity of many PCR reactions in various applications. Touchdown PCR uses an annealing/extension temperature that is higher than the T_m of the primers during the initial PCR cycles. The annealing/extension temperature is then reduced to the primer T_m for the remaining PCR cycles. The change can be performed in a single step or in increments over several cycles.
Negative control shows a PCR product or “false positive” results are obtained.	Reagents are contaminated.	Sigma recommends that a reagent blank without DNA template be included as a control in every PCR run to determine if the reagents used in extraction or PCR are contaminated with a template from a previous reaction.

AlumaSeal is a trademark of Excel Scientific, Inc.

JumpStart and JumpStart *Taq* Antibody are licensed under U.S. Patent No. 5,338,671 and 5,587,287 and corresponding patents in other countries.

CAK/DJD/MAM 4/03

Sigma brand products are sold through Sigma-Aldrich, Inc.

Sigma-Aldrich, Inc. warrants that its products conform to the information contained in this and other Sigma-Aldrich publications.

Purchaser must determine the suitability of the product(s) for their particular use. Additional terms and conditions may apply.

Please see reverse side of the invoice or packing slip.